



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/089,997	09/23/2002	Barrie Hayes-Gill	469.1094	5818
21171	7590	03/07/2006	EXAMINER	
STAAS & HALSEY LLP SUITE 700 1201 NEW YORK AVENUE, N.W. WASHINGTON, DC 20005			KRAMER, NICOLE R	
			ART UNIT	PAPER NUMBER
			3762	

DATE MAILED: 03/07/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/089,997

Applicant(s)

HAYES-GILL ET AL.

Examiner

Nicole R. Kramer

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 12 December 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 25-62 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☐ Claim(s) _____ is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☒ Claim(s) 25-62 are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Election/Restrictions

1. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 25-47, drawn to an apparatus for determining the heart rate of a fetus, including a processor which determines the heart rate of the mother.

Group II, claim(s) 48-62, drawn to a method for determining the heart rate of the fetus, including positioning a detector in accordance with a determined position of the fetus.

2. The inventions listed as Groups I and II do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

With respect to Group I (and independent claim 25), the prior art of reference (U.S. Patent No. 4,781,200 to Baker) teaches an apparatus for detecting the heart rate of a fetus (fetal monitoring system 20), including at least two electrodes positioned on the abdomen of the mother in use for detecting ECG signals (plurality of fetal monitoring sensors 35 may be ECG sensors; see col. 5, lines 1-5), a processor for receiving the signals received from each detector and determining the heart rate of the fetus (control unit 40; see col. 7, lines 20-35). Examiner thus considers the special technical feature

Art Unit: 3762

of claim 25 to be that the processor also includes means for determining the heart rate of the mother.

With respect to claim Group II (and independent claim 48), the prior art of reference (U.S. Patent No. 4,781,200 to Baker) teaches a method of determining the heart rate of the fetus, including the steps of placing the ECG detector on the abdomen of the mother, monitoring the ECG signals for a predetermined amount of time, and processing the obtained ECG signals in order to determine the heart rate of the fetus. Examiner thus considers the special technical feature of claim 48 to be positioning the ECG detector in accordance with a determined position of the fetus.

Since the two above-listed Groups do not share a special technical feature, restriction is proper.

3. As to the time for making the present restriction requirement, Examiner notes that a restriction requirement is normally made before an action upon the merits. However, it may be made at any time before final action in the case at the discretion of the examiner (see MPEP Section 811). Although restriction was proper before the previous action upon the merits, Examiner did not make the restriction requirement at the time because the relatively short amount of claims did not impose a serious burden on the Examiner. However, in light of the amount of newly added claims, the restriction requirement is considered to be necessary in order to avoid a serious burden on the Examiner. Applicant is invited to contact the Examiner at the below contact information with any questions or concerns.

4. Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.


5. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).


Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Nicole R. Kramer whose telephone number is 571-272-8792. The examiner can normally be reached on Monday through Friday, 8 a.m. to 4:30 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 571-272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).


NRK
3/2/06


George Manuel
Primary Examiner